

HOMEWORK #5
Due at the start of Class on Thursday 11/09/06

Readings:

Review Chapters 10 and 11;

Problems:

1. You have been asked to design an ultrasound system for imaging of the heart. The system must be capable of acquiring 40 frames a second at a maximum depth of 20 cm.
 - a) Determine how many lines per frame can be acquired. Assume that the speed of sound is 1500 m/s.
 - b) Determine the highest frequency that can be used in order that the waves not be attenuated by more than 99%. Assume an attenuation of 1dB/cm/MHz.
 - c) Determine the size of the detector such that the entire field of view will be in the near field. Use the frequency derived in part b.
 - d) Determine the depth resolution, assuming that the temporal pulse duration is equal to 3 cycles of the acoustic wave.

2. Consider a transducer of dimensions $L \times L$ operating at a frequency of 3 MHz.
 - a) Determine the size L of the transducer such that the far field region begins at 30 cm.
 - b) Sketch the 2D far field pattern as a function of z .
 - c) Consider two point reflectors at $(d/2, 0, z)$ and $(-d/2, 0, z)$. If the resolution is defined as the effective width of the field pattern, determine the minimum distance d between the two points such that the two points can still be resolved. In other words, the distance should be equal to the effective width of the field pattern.
 - d) Now assume that an acoustic lens has been added to the transducer to focus the beam at a focal depth of 15 cm. What is the minimum separation of points that can be resolved at the focal depth?

3. Problem 11.12

4. Problem 11.14